

Issues

#	Project	Tracker	Status	Priority	Subject	Assignee	Updated	Target version	Estimated time	% Done
216	FC005-ElectronicLoad	Anomalie	Résolu	Normal	Add a second fan to balance current between -12/+12 and for better heat protection		07/07/2020 12:11 AM	1.1	2.00	100
215	FC005-ElectronicLoad	Anomalie	Résolu	Normal	Security fan is currently progressively started		07/07/2020 12:13 AM	1.1	1.00	100